

### **INTENDED USE**

This human coagulation Factor IX antigen assay is intended for the quantitative determination of total Factor IX antigen in human plasma. **For research use only.**

### **BACKGROUND**

Factor IX (aka Christmas Factor) is a single-chain, 415 amino acid glycoprotein zymogens [1]. Factor IX is activated by either Factor XIa or the Factor VIIa complex [2]. Factor IXa converts Factor X to Factor Xa during the intrinsic pathway of the coagulation cascade. Factor IX is used to treat patients with hemophilia B, an X-linked bleeding disorder [3].

### **ASSAY PRINCIPLE**

Human Factor IX will bind to the affinity purified capture antibody coated on the microtiter plate. Factor IX and IXa will react with the antibody on the plate. After appropriate washing steps, polyclonal anti-human Factor IX primary antibody binds to the captured protein. Excess primary antibody is washed away and bound antibody, which is proportional to the total Factor IX present in the samples, is reacted with the secondary antibody. Following an additional washing step, TMB substrate is used for color development at 450nm. A standard calibration curve is prepared along with the samples to be measured using dilutions of human Factor IX. Color development is proportional to the concentration of Factor IX in the samples.

### **STANDARD CALIBRATION**

Factor IX standard provided is calibrated against the WHO 4th International Standard for Factor IX, Plasma, Human distributed by NIBSC (09/172), South Mimms, Potters Bar, Hertfordshire, UK. **See C of A for lot specific Calibration of Standard.**

### **REAGENTS PROVIDED**

- **96-well antibody coated microtiter strip plate** (removable wells 8x12) containing anti-human Factor IX antibody, blocked and dried.
- **10X Wash buffer:** 1 bottle of 50ml
- **Human Factor IX standard:** 1 vial lyophilized standard
- **Anti-human Factor IX primary antibody:** 1 vial lyophilized polyclonal antibody
- **Anti-goat horseradish peroxidase secondary antibody:** 1 vial concentrated HRP labeled antibody
- **TMB substrate solution:** 1 bottle of 10ml solution

### **STORAGE AND STABILITY**

**The secondary antibody must be stored at  $\leq -70^{\circ}\text{C}$ .** Store all other kit components at  $4^{\circ}\text{C}$  upon arrival. Return any unused microplate strips to the plate pouch with desiccant. Reconstituted standard and primary may be stored at  $-80^{\circ}\text{C}$  for later use. Do not freeze-thaw the standard and primary antibody more than once. Store all other unused kit components at  $4^{\circ}\text{C}$ . This kit should not be used beyond the expiration date.

### **OTHER REAGENTS AND SUPPLIES REQUIRED**

- Microtiter plate shaker capable of 300 rpm uniform horizontally circular movement
- Manifold dispenser/aspirator or automated microplate washer
- Microplate reader capable of measuring absorbance at 450 nm
- Pipettes and Pipette tips
- Deionized or distilled water
- Polypropylene tubes for dilution of standard
- Paper towels or laboratory wipes
- 1N  $\text{H}_2\text{SO}_4$  or 1N HCl
- Bovine Serum Albumin Fraction V (BSA)
- Tris(hydroxymethyl)aminomethane (Tris)
- Sodium Chloride (NaCl)

## PRECAUTIONS

- FOR LABORATORY RESEARCH USE ONLY. NOT FOR DIAGNOSTIC USE.
- Do not mix any reagents or components of this kit with any reagents or components of any other kit. This kit is designed to work properly as provided.
- Always pour peroxidase substrate out of the bottle into a clean test tube. Do not pipette out of the bottle as contamination could result.
- Keep plate covered except when adding reagents, washing, or reading.
- DO NOT pipette reagents by mouth and avoid contact of reagents and specimens with skin.
- DO NOT smoke, drink, or eat in areas where specimens or reagents are being handled.

## PREPARATION OF REAGENTS

- **TBS buffer:** 0.1M Tris, 0.15M NaCl, pH 7.4
- **Blocking buffer (BB):** 3% BSA (w/v) in TBS
- **1X Wash buffer:** Dilute 50ml of 10X wash buffer concentrate with 450ml of deionized water

## SAMPLE COLLECTION

Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000xg within 30 minutes of collection. Assay immediately or aliquot and store at  $\leq -20^{\circ}\text{C}$ . Avoid repeated freeze-thaw cycles.

## ASSAY PROCEDURE

Perform assay at room temperature. Vigorously shake plate (300rpm) at each step of the assay.

### Preparation of Standard

Reconstitute standard according to the **C of A**. Agitate gently to completely dissolve contents.

Dilution table for preparation of human Factor IX standard:

Factor IX concentration (ng/ml)	Dilutions
100	900 $\mu\text{L}$ (BB) + 100 $\mu\text{L}$ standard from vial
50	500 $\mu\text{L}$ (BB) + 500 $\mu\text{L}$ (100ng/mL)
25	500 $\mu\text{L}$ (BB) + 500 $\mu\text{L}$ (50ng/mL)
10	600 $\mu\text{L}$ (BB) + 400 $\mu\text{L}$ (25ng/mL)
5	500 $\mu\text{L}$ (BB) + 500 $\mu\text{L}$ (10ng/mL)
2.5	500 $\mu\text{L}$ (BB) + 500 $\mu\text{L}$ (5ng/mL)
1	600 $\mu\text{L}$ (BB) + 400 $\mu\text{L}$ (2.5ng/mL)
0.5	500 $\mu\text{L}$ (BB) + 500 $\mu\text{L}$ (1ng/mL)
0.25	500 $\mu\text{L}$ (BB) + 500 $\mu\text{L}$ (0.5ng/mL)
0.1	600 $\mu\text{L}$ (BB) + 400 $\mu\text{L}$ (0.25ng/mL)
0	500 $\mu\text{L}$ (BB) Zero point to determine background

**NOTE: DILUTIONS FOR THE STANDARD CURVE AND ZERO STANDARD MUST BE MADE AND APPLIED TO THE PLATE IMMEDIATELY.**

### Standard and Unknown Addition

Remove microtiter plate from bag and add 100 $\mu\text{L}$  Factor IX standards (in duplicate) and unknowns to wells. Carefully record position of standards and unknowns. Shake plate at 300rpm for 30 minutes. Wash wells three times with 300 $\mu\text{L}$  wash buffer. Remove excess wash by gently tapping plate on paper towel or kimwipe.

NOTE: The assay measures Factor IX antigen in the 0.1-100 ng/ml range. If the unknown is thought to have high Factor IX levels, dilutions may be made in blocking buffer. A 1:1,000 to 1:4,000 dilution for normal plasma is suggested for best results.

### Primary Antibody Addition

Reconstitute primary antibody by adding 10ml of blocking buffer directly to the vial and agitate gently to completely dissolve contents. Add 100 $\mu\text{L}$  to all wells. Shake plate at 300rpm for 30 minutes. Wash wells three times with 300 $\mu\text{L}$  wash buffer. Remove excess wash by gently tapping plate on paper towel or kimwipe.

### Secondary Antibody Addition

Briefly centrifuge vial before opening. **See C of A for lot specific dilution instructions.** Add 100 $\mu\text{L}$  of your dilution to all wells. Shake plate at 300rpm for 30 minutes. Wash wells three times with 300 $\mu\text{L}$  wash buffer. Remove excess wash by gently tapping plate on paper towel or kimwipe.

### Substrate Incubation

Add 100µl TMB substrate to all wells and shake plate for 2-10 minutes. Substrate will change from colorless to different strengths of blue. Quench reaction by adding 50µl of 1N H<sub>2</sub>SO<sub>4</sub> or HCl stop solution to all wells when samples are visually in the same range as the standards. Add stop solution to wells in the same order as substrate upon which color will change from blue to yellow. Mix thoroughly by gently shaking the plate.



### Measurement

Set the absorbance at 450nm in a microtiter plate spectrophotometer. Measure the absorbance in all wells at 450nm. Subtract zero point from all standards and unknowns to determine corrected absorbance (A<sub>450</sub>).

### Calculation of Results

Plot A<sub>450</sub> against the amount of Factor IX in the standards. Fit a straight line through the linear points of the standard curve using a linear fit procedure if unknowns appear on the linear portion of the standard curve. Alternatively, create a standard curve by analyzing the data using a software program capable of generating a four-parameter logistic (4PL) curve fit. The amount of Factor IX in the unknowns can be determined from this curve. If samples have been diluted, the calculated concentration must be multiplied by the dilution factor.

A typical standard curve (EXAMPLE ONLY):



### EXPECTED VALUES

The concentration of Factor IX in normal human plasma is about 5 µg/mL [4].

### PERFORMANCE CHARACTERISTICS

**Sensitivity:** The minimum detectable dose (MDD) was determined by adding two standard deviations to the mean optical density value of twenty zero standard replicates and calculating the corresponding concentration. **See C of A for lot specifications.**

**Intra-assay Precision:** Three samples of known concentration were tested twenty times on one plate to assess intra-assay precision.

Sample	1	2	3
n	20	20	20
Mean (ng/ml)	0.226	1.56	5.33
Standard Deviation	0.02	0.095	0.362
CV (%)	8.84	6.09	6.8

**Inter-assay Precision:** Three samples of known concentration were tested in ten independent assays to assess inter-assay precision.

Sample	1	2	3
n	10	10	10
Mean (ng/ml)	0.811	3.07	16.2
Standard Deviation	0.064	0.133	1.19
CV (%)	7.93	4.34	7.38

**Recovery:** The recovery of antigen spiked to levels throughout the range of the assay in diluted plasma was evaluated.

Sample	1	2	3	4
n	4	4	4	4
Mean (ng/ml)	0.4	1.95	8.58	16.2
Average % Recovery	100	98	114	108
Range	87-119%	89-113%	107-119%	106-111%

**Linearity:** To assess the linearity of the assay, pooled citrated human plasma samples containing high concentrations of antigen were serially diluted to produce samples with values within the dynamic range of the assay.

Sample	1:2	1:4	1:8	1:16
n	4	4	4	4
Average % of Expected	100	91	91	94
Range	94-103%	89-94%	88-93%	91-98%

**Specificity:** This assay recognizes total human Factor IX. Pooled normal plasma from mouse, rat, pig, dog, cyno monkey, and rhesus monkey were assayed and no significant cross-reactivity was observed.

**DISCLAIMER**

This information is believed to be correct but does not claim to be all-inclusive and should be used only as a guide. The supplier of this kit shall not be held liable for any damage resulting from handling of or contact with the above product.

**REFERENCES**

1. DiScipio RG, *et al.*: Biochemistry. 1977, 16(4):698-706.
2. Soulier JP, Wartelle O, and Ménaché D: Rev Franc Etud Clin Biol. 1958, 3:263-267.
3. Roth DA, *et al.*: Blood. 2001, 98(13):3600-3606.
4. Thompson AR: Blood. 1986, 967(3):565-572.

**Example of ELISA Plate Layout**

**96 Well Plate: 22 Standard wells, 74 Sample wells**

	1	2	3	4	5	6	7	8	9	10	11	12
A	0	0.1 ng/ml	0.25 ng/ml	0.5 ng/ml	1 ng/ml	2 ng/ml	5 ng/ml	10 ng/ml	25 ng/ml	50 ng/ml	100 ng/ml	
B	0	0.1 ng/ml	0.25 ng/ml	0.5 ng/ml	1 ng/ml	2 ng/ml	5 ng/ml	10 ng/ml	25 ng/ml	50 ng/ml	100 ng/ml	
C												
D												
E												
F												
G												
H												

